



Certificate No.L0372-131125

財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

This is to certify that

TUNG PEI INDUSTRIAL CO., LTD.

Length Calibration Laboratory

7, Sung-Chiang N. Rd., Chung-Li Ind. Zone, Taoyuan Hsien 320, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2005
Accreditation Number : 0372
Originally Accredited : January 01, 1999
Effective Period : January 01, 2014 to December 31, 2016
Accredited Scope : Calibration Field, see described in the Appendix

Jay-San Chen
President, Taiwan Accreditation Foundation
Date : November 25, 2013



Certificate No.L0372-131125

財團法人全國認證基金會
Taiwan Accreditation Foundation

Accreditation Number : 0372

Laboratory Head : LIN, Ho-Nein

Length

Calibration items	Calibration Method /Working Standard	Ranges	Smallest uncertainty
KA1001 Gauge Block (CNS 8092 /112 pcs) (Steel) (Null below)	Q10IT901A /P.T.W B-1 Grade 00	(0.5 to 100) mm	$[(92)^2+(1.8L)^2]^{1/2}$ nm L:mm, length of gauge block

Remarks : Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.

Approval Signatory	Scope
LIN, Ho-Nein	KA1001
KAO, Shun-Liang	KA1001